

CERTIFICATE

Number: C – 957403-1

Date: 25/07/2017 Budapest

Type and identification

data of product: Delta XX-T-S-E-V, chemical fume hood series

Model variants identification details:

XX- size designation, External width (dm): 10, 11, 12, 13, 15, 19, 21

T- type: B- Basic, D- distillation, W- walk-in, T- transfer

S- surface: N- normal HPL laminated, PP- Polypropylene cover, ESD- electrostatic discharge

E- suction power modes: O- constant power (100%),
E- economic, Two power levels 30% or 100%

V- VAV variable air volume, Continuous suction power 30%-100%

Produced by: FOTON Kft./ Hungary

Client: FOTON Kft. H-2120 Dunakeszi, Táltos u. 28.

On the basis of the documents submitted by the client and the test results (R-957403) under registration number, ÉMI TÜV SÜD Kft. certifies that the samples

complies

with the below listed legal rules, relevant stipulations, standards and/or attested quality.

Applied directives, regulations, standards:

MSZ EN 14175-1:2003, MSZ EN 14175-2:2003, MSZ EN 14175-3:2004
MSZ EN 14175-6:2006, MSZ EN 14175-7:2012

Valid until: 2 years from the issue, or the change in product, production, production technology and the concerning regulations



Gabriella Süvegesné Váradi
 Gabriella Süvegesné Váradi
 Head of department
 ÉMI-TÜV SÜD Kft.



The CE mark should be showed on the product compliance with the adequate technical documentation and EC conformity declaration observing all the relevant laws.

In accordance with the above mentioned directive the applicant has to inform us about any modification made or planned to be made on the above mentioned product. This certificate is issued under the contract of 957403.

TANÚSÍTVÁNY

Száma: C-957403-1

Kelt: Budapest, 2017. 07. 25.

Termékcsoport neve: **Delta XX-T-S-E-V, vegyifülke sorozat**

Típusváltozatok azonosító adatai:

XX- méret jelölés, külső szélesség (dm): 10, 11, 12, 13, 15, 19, 21

T- típus: B- normál, D- desztilláló, W- bejáró, T- átadó

S- belső burkolat : N- normal HPL laminált, PP- Polypropilén borítás,
ESD- electrostatikus borítás

E- szívóteljesítmény üzemmódok: O- állandó teljesítmény (100%),
E- gazdaságos, két teljesítmény 30% vagy 100

V- VAV változó légmennyiség, folyamatos szívóteljesítmény 30%-100%

Megbízó/ gyártó: **FOTON Kft.**
2120 Dunakeszi, Táltos u. 28.

Az ÉMI-TÜV SÜD Kft. KERMI Osztály a kérelmező által benyújtott dokumentációk és a vizsgálati eredmények (R-957403) alapján megállapítja, hogy a vizsgált minták az alábbiakban részletezett jogszabályoknak, vonatkozó előírásoknak

megfelel

Alkalmazott, előírások, szabványok:

MSZ EN 14175-1:2003, MSZ EN 14175-2:2003, MSZ EN 14175-3:2004

MSZ EN 14175-6:2006, MSZ EN 14175-7:2012

Érvényes: A kiállítástól számított 2 évig, illetve a termékben, valamint a vonatkozó jogszabályokban, előírásokban bekövetkező változásokig.



Ph.

Süvegné Váradi Gabriella
Süvegné Váradi Gabriella
osztályvezető
ÉMI-TÜV SÜD Kft.



A terméken a CE jelölés elhelyezendő a megfelelő műszaki dokumentáció és az EK megfelelési igazolás kiállítása illetve az összes vonatkozó jogszabályok betartása mellett.

A megbízó köteles értesíteni a megbízottat a termékre vonatkozó minden tervezett, vagy bekövetkezett változásokról. Jelen tanúsítvány a 957403 számú szerződés alapján lett kiállítva.

ÉMI-TÜV SÜD Kft.

H-1043 Budapest, Dugonics u. 11.

Tel: +36 1 210-9570

Fax: +36 1 314-3820

ÉMI-TÜV SÜD Ltd.